

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091622\
 Data File : VN074518.D
 Acq On : 16 Sep 2022 18:26
 Operator : JC\MD
 Sample : N4628-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Sep 16 22:35:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	319882	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.904	114	630234	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	623556	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	273532	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	260397	47.977	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.960%
35) Dibromofluoromethane	7.957	113	211872	50.608	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.220%
50) Toluene-d8	10.381	98	779159	48.123	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.240%
62) 4-Bromofluorobenzene	12.675	95	298556	53.839	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.680%
Target Compounds						
16) Acetone	4.228	43	10063	2.897	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091622\
Data File : VN074518.D
Acq On : 16 Sep 2022 18:26
Operator : JC\MD
Sample : N4628-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Sep 16 22:35:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

