

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091622\
 Data File : VN074530.D
 Acq On : 17 Sep 2022 00:06
 Operator : JC\MD
 Sample : N4549-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-47

Quant Time: Sep 17 06:17:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	300354	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	580306	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	584985	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	258986	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	241687	47.425	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.840%
35) Dibromofluoromethane	7.951	113	194151	50.365	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.740%
50) Toluene-d8	10.380	98	737110	49.443	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.880%
62) 4-Bromofluorobenzene	12.674	95	281475	55.126	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.260%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.281	59	61284	36.915	ug/l #	89
16) Acetone	4.228	43	15291	4.689	ug/l #	67
19) Methyl tert-butyl Ether	5.528	73	76847	4.609	ug/l	100
22) Diisopropyl ether	6.416	45	16509	0.870	ug/l #	70
31) Cyclohexane	8.028	56	13536	1.496	ug/l #	63
73) Isopropylbenzene	12.527	105	6345	0.239	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091622\
Data File : VN074530.D
Acq On : 17 Sep 2022 00:06
Operator : JC\MD
Sample : N4549-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-47

Quant Time: Sep 17 06:17:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

