

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091624\
 Data File : VN083907.D
 Acq On : 16 Sep 2024 09:45
 Operator : JC\MD
 Sample : VN0916MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0916MBL01

Quant Time: Sep 17 02:27:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	148161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	248574	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	209075	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	85036	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	114070	51.060	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.120%
35) Dibromofluoromethane	8.165	113	80408	50.529	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.060%
50) Toluene-d8	10.565	98	297873	54.145	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.300%
62) 4-Bromofluorobenzene	12.847	95	99529	43.851	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091624\
Data File : VN083907.D
Acq On : 16 Sep 2024 09:45
Operator : JC\MD
Sample : VN0916MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0916MBL01

Quant Time: Sep 17 02:27:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

