

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091624\
 Data File : VN083921.D
 Acq On : 16 Sep 2024 16:39
 Operator : JC\MD
 Sample : P3886-01DL 400X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-06-6.5-090624DL

Quant Time: Sep 17 02:34:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	170976	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	290261	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	245572	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	104988	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	128990	50.033	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.060%	
35) Dibromofluoromethane	8.171	113	91748	49.374	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.740%	
50) Toluene-d8	10.565	98	343811	53.520	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.040%	
62) 4-Bromofluorobenzene	12.847	95	122855	46.354	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 92.700%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.377	101	17998	9.191	ug/l	98
27) cis-1,2-Dichloroethene	7.489	96	2651	1.127	ug/l	89
44) Trichloroethene	9.353	130	107840	55.318	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091624\
Data File : VN083921.D
Acq On : 16 Sep 2024 16:39
Operator : JC\MD
Sample : P3886-01DL 400X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-06-6.5-090624DL

Quant Time: Sep 17 02:34:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

