

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091624\
 Data File : VN083923.D
 Acq On : 16 Sep 2024 17:27
 Operator : JC\MD
 Sample : P3876-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-SOIL-REF-6

Quant Time: Sep 17 02:35:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	163719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	282777	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	238124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101809	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	125750	50.939	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.880%
35) Dibromofluoromethane	8.171	113	87916	48.564	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.120%
50) Toluene-d8	10.565	98	333922	53.357	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.720%
62) 4-Bromofluorobenzene	12.847	95	124337	48.155	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091624\
Data File : VN083923.D
Acq On : 16 Sep 2024 17:27
Operator : JC\MD
Sample : P3876-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-SOIL-REF-6

Quant Time: Sep 17 02:35:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

