

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091625\
 Data File : VN087848.D
 Acq On : 16 Sep 2025 14:43
 Operator : JC\MD
 Sample : VN0916MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0916MBL01

Quant Time: Sep 17 01:20:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	189805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	429890	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	399187	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	183370	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	191021	49.120	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.240%
35) Dibromofluoromethane	8.147	113	152947	49.277	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.560%
50) Toluene-d8	10.547	98	548510	47.216	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.440%
62) 4-Bromofluorobenzene	12.829	95	182318	44.130	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091625\
Data File : VN087848.D
Acq On : 16 Sep 2025 14:43
Operator : JC\MD
Sample : VN0916MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0916MBL01

Quant Time: Sep 17 01:20:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

