

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091625\
 Data File : VN087861.D
 Acq On : 16 Sep 2025 19:19
 Operator : JC\MD
 Sample : Q3021-05 200X
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 821-I

Quant Time: Sep 17 01:29:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	257495	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	579892	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	559964	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	283351	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	250255	47.435	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.860%
35) Dibromofluoromethane	8.153	113	201206	48.057	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.120%
50) Toluene-d8	10.547	98	732504	46.744	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.480%
62) 4-Bromofluorobenzene	12.829	95	279549	50.162	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091625\
Data File : VN087861.D
Acq On : 16 Sep 2025 19:19
Operator : JC\MD
Sample : Q3021-05 200X
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
821-I

Quant Time: Sep 17 01:29:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

