

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091625\
 Data File : VN087863.D
 Acq On : 16 Sep 2025 20:01
 Operator : JC\MD
 Sample : Q3087-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAGEMENT

Quant Time: Sep 17 01:30:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 02:55:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	236522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	523081	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	508812	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	261215	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	233622	48.209	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.420%
35) Dibromofluoromethane	8.147	113	184824	48.939	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.880%
50) Toluene-d8	10.547	98	666821	47.174	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.340%
62) 4-Bromofluorobenzene	12.829	95	248887	49.511	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091625\
Data File : VN087863.D
Acq On : 16 Sep 2025 20:01
Operator : JC\MD
Sample : Q3087-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAGEMENT

Quant Time: Sep 17 01:30:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 02:55:42 2025
Response via : Initial Calibration

