

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091718\
 Data File : VN051279.D
 Acq On : 17 Sep 2018 19:02
 Operator : MD\SY
 Sample : J4929-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-PIT-6

Quant Time: Sep 19 10:06:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 18 06:04:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	450208	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	667383	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	561157	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	187873	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	242297	49.52	ug/l	0.00
Spiked Amount						
			Recovery	=	99.04%	
35) Dibromofluoromethane	7.59	113	258594	49.86	ug/l	0.00
Spiked Amount						
			Recovery	=	99.72%	
50) Toluene-d8	10.09	98	877963	45.80	ug/l	0.00
Spiked Amount						
			Recovery	=	91.60%	
62) 4-Bromofluorobenzene	12.40	95	241596	39.62	ug/l	0.00
Spiked Amount						
			Recovery	=	79.24%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	13485	2.51	ug/l	93
20) Methylene Chloride	4.55	84	90493	16.90	ug/l	99
25) 2-Butanone	6.85	43	5620	3.59	ug/l	94
43) Isopropyl Acetate	8.17	43	171468	26.68	ug/l	98
59) 2-Hexanone	10.77	43	12247	5.61	ug/l #	59
95) Naphthalene	15.13	128	162358	30.27	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091718\
Data File : VN051279.D
Acq On : 17 Sep 2018 19:02
Operator : MD\SY
Sample : J4929-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-PIT-6

Quant Time: Sep 19 10:06:26 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 18 06:04:06 2018
Response via : Initial Calibration

