

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091719\
 Data File : VN058136.D
 Acq On : 17 Sep 2019 10:15
 Operator : JC/SP
 Sample : PB123118TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB123118TB

Quant Time: Sep 18 07:44:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	358750	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	590180	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	552964	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	291421	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	350590	48.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.50%	
35) Dibromofluoromethane	7.57	113	243567	44.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.42%	
50) Toluene-d8	10.08	98	958429	45.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.64%	
62) 4-Bromofluorobenzene	12.40	95	324411	40.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.50%	
Target Compounds						
18) Methyl Acetate	4.30	43	38411	6.048	ug/l #	78
43) Isopropyl Acetate	8.15	43	124906	10.503	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091719\
Data File : VN058136.D
Acq On : 17 Sep 2019 10:15
Operator : JC/SP
Sample : PB123118TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB123118TB

Quant Time: Sep 18 07:44:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

