

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091719\
 Data File : VN058138.D
 Acq On : 17 Sep 2019 10:59
 Operator : JC/SP
 Sample : K4871-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VE-1-4

Quant Time: Sep 18 07:50:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	289174	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	476139	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	429669	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	235505	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	342255	58.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.88%	
35) Dibromofluoromethane	7.57	113	239823	54.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.14%	
50) Toluene-d8	10.08	98	912578	53.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.98%	
62) 4-Bromofluorobenzene	12.40	95	310676	47.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.56%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091719\
Data File : VN058138.D
Acq On : 17 Sep 2019 10:59
Operator : JC/SP
Sample : K4871-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VE-1-4

Quant Time: Sep 18 07:50:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

