

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091719\
 Data File : VN058149.D
 Acq On : 17 Sep 2019 15:02
 Operator : JC/SP
 Sample : K4922-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GRAB-40150

Manual Integrations
 APPROVED

MMDadoda
 9/18/2019 4:41:46 PM

Quant Time: Sep 18 08:26:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 10 03:02:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	351713	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	580304	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	528261	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	294083	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	332739	46.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.42%	
35) Dibromofluoromethane	7.58	113	235804	44.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.06%	
50) Toluene-d8	10.08	98	909715	43.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.50%	
62) 4-Bromofluorobenzene	12.41	95	323443	40.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.62%	
Target Compounds						
11) Tert butyl alcohol	4.76	59	21607m	25.061	ug/l	Qvalue
16) Acetone	3.80	43	44004	18.843	ug/l	98
19) Methyl tert-butyl Ether	5.03	73	84098	5.396	ug/l	98
22) Diisopropyl ether	5.93	45	85758	4.989	ug/l	90
29) Tetrahydrofuran	7.20	42	32596	15.651	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091719\
Data File : VN058149.D
Acq On : 17 Sep 2019 15:02
Operator : JC/SP
Sample : K4922-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GRAB-40150

Manual Integrations
APPROVED

MMDadoda
9/18/2019 4:41:46 PM

Quant Time: Sep 18 08:26:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 10 03:02:13 2019
Response via : Initial Calibration

