


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091719\
 Data File : VN058133.D
 Acq On : 17 Sep 2019 8:19
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091019W.M
 Title : SW846 8260
 Last Update : Tue Sep 10 03:02:13 2019



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.9	23562	PASS
75	95	30	60	55.6	59856	PASS
95	95	100	100	100.0	107674	PASS
96	95	5	9	7.0	7514	PASS
173	174	0.00	2	0.4	329	PASS
174	95	50	100	74.4	80160	PASS
175	174	5	9	7.3	5846	PASS
176	174	95	101	98.9	79266	PASS
177	176	5	9	6.7	5303	PASS