

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091720\
 Data File : VN063419.D
 Acq On : 17 Sep 2020 15:14
 Operator : JC/MD
 Sample : PB131674ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131674ZHE#02

Quant Time: Sep 18 03:50:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	240381	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	456052	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	458913	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	194381	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	197685	48.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.62%	
35) Dibromofluoromethane	7.55	113	142774	45.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.44%	
50) Toluene-d8	10.06	98	582376	49.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.12%	
62) 4-Bromofluorobenzene	12.38	95	231545	53.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.42%	
Target Compounds						
16) Acetone	3.78	43	13792	10.488	ug/l	95
18) Methyl Acetate	4.29	43	5798	1.737	ug/l #	72
20) Methylene Chloride	4.51	84	14512	4.355	ug/l	96
43) Isopropyl Acetate	8.13	43	89973	10.552	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091720\
Data File : VN063419.D
Acq On : 17 Sep 2020 15:14
Operator : JC/MD
Sample : PB131674ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131674ZHE#02

Quant Time: Sep 18 03:50:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

