

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091720\
 Data File : VN063421.D
 Acq On : 17 Sep 2020 16:01
 Operator : JC/MD
 Sample : PB131677TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB131677TB

Manual Integrations
 APPROVED

MMDadoda
 9/18/2020 2:45:06 PM

Quant Time: Sep 18 03:54:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	246787	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	466500	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	471987	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	203456	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	201522	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
35) Dibromofluoromethane	7.55	113	146886	45.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.96%	
50) Toluene-d8	10.06	98	610715	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
62) 4-Bromofluorobenzene	12.38	95	241807	54.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.64%	
Target Compounds						
16) Acetone	3.78	43	12648	9.368	ug/l	95
18) Methyl Acetate	4.29	43	5330m	1.555	ug/l	
20) Methylene Chloride	4.51	84	8984	2.626	ug/l #	83
43) Isopropyl Acetate	8.13	43	97933	11.228	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091720\
Data File : VN063421.D
Acq On : 17 Sep 2020 16:01
Operator : JC/MD
Sample : PB131677TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB131677TB

Manual Integrations
APPROVED

MMDadoda
9/18/2020 2:45:06 PM

Quant Time: Sep 18 03:54:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

