

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091720\
 Data File : VN063422.D
 Acq On : 17 Sep 2020 16:25
 Operator : JC/MD
 Sample : L4038-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAV-091620-DP

Manual Integrations
 APPROVED

MMDadoda
 9/18/2020 2:45:09 PM

Quant Time: Sep 18 03:57:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	237379	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	451058	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	448283	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	196720	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	196898	48.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.46%	
35) Dibromofluoromethane	7.55	113	146726	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
50) Toluene-d8	10.06	98	583662	49.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.42%	
62) 4-Bromofluorobenzene	12.38	95	232940	54.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.24%	
Target Compounds						
16) Acetone	3.78	43	33187	25.555	ug/l	98
18) Methyl Acetate	4.29	43	4253m	1.290	ug/l	
20) Methylene Chloride	4.51	84	17810	5.413	ug/l	92
43) Isopropyl Acetate	8.13	43	95831	11.364	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091720\
Data File : VN063422.D
Acq On : 17 Sep 2020 16:25
Operator : JC/MD
Sample : L4038-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAV-091620-DP

Manual Integrations
APPROVED

MMDadoda
9/18/2020 2:45:09 PM

Quant Time: Sep 18 03:57:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

