

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091720\
 Data File : VN063427.D
 Acq On : 17 Sep 2020 18:25
 Operator : JC/MD
 Sample : L4015-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REACTOR-5W-(D-0)

Quant Time: Sep 18 04:06:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	225628	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	429002	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	424993	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	181855	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	189649	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
35) Dibromofluoromethane	7.55	113	136959	46.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.24%	
50) Toluene-d8	10.06	98	553710	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
62) 4-Bromofluorobenzene	12.38	95	220291	53.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.62%	
Target Compounds						
16) Acetone	3.78	43	14492	11.740	ug/l	95
18) Methyl Acetate	4.29	43	5407	1.725	ug/l #	87
20) Methylene Chloride	4.51	84	14198	4.540	ug/l	90
43) Isopropyl Acetate	8.13	43	75286	9.386	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091720\
Data File : VN063427.D
Acq On : 17 Sep 2020 18:25
Operator : JC/MD
Sample : L4015-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REACTOR-5W-(D-0)

Quant Time: Sep 18 04:06:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

