

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091720\
 Data File : VN063439.D
 Acq On : 17 Sep 2020 23:59
 Operator : JC/MD
 Sample : L4021-02
 Misc : 10.47µ/5mL/100µL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B14-5-7

Quant Time: Sep 26 02:40:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	222189	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	410488	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	406881	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	182198	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	179985	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
35) Dibromofluoromethane	7.55	113	129992	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	
50) Toluene-d8	10.06	98	521759	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
62) 4-Bromofluorobenzene	12.38	95	209331	53.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.88%	
Target Compounds						
18) Methyl Acetate	4.28	43	52104	16.885	ug/l	Qvalue 100
44) Trichloroethene	8.80	130	8417	2.780	ug/l	99
64) Tetrachloroethene	10.60	164	286436	103.385	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091720\
Data File : VN063439.D
Acq On : 17 Sep 2020 23:59
Operator : JC/MD
Sample : L4021-02
Misc : 10.47µL/5mL/100µL/5.00mL/MSVOA_N/MEOH
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B14-5-7

Quant Time: Sep 26 02:40:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

