

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN091720\
 Data File : VN063442.D
 Acq On : 18 Sep 2020 1:10
 Operator : JC/MD
 Sample : L4021-05
 Misc : 10.61g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B15A-5-7

Quant Time: Sep 18 05:29:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	234466	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	437884	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	434530	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	195572	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	193208	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
35) Dibromofluoromethane	7.55	113	138638	45.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.46%	
50) Toluene-d8	10.06	98	559619	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
62) 4-Bromofluorobenzene	12.38	95	222867	53.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.68%	
Target Compounds						
18) Methyl Acetate	4.28	43	52206	16.032	ug/l	98
27) cis-1,2-Dichloroethene	6.78	96	7479	2.254	ug/l	94
44) Trichloroethene	8.80	130	20538	6.359	ug/l	100
64) Tetrachloroethene	10.60	164	343814	116.199	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091720\
Data File : VN063442.D
Acq On : 18 Sep 2020 1:10
Operator : JC/MD
Sample : L4021-05
Misc : 10.61µL/5mL/100µL/5.00mL/MSVOA_N/MEOH
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B15A-5-7

Quant Time: Sep 18 05:29:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

