

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091720\
 Data File : VN063443.D
 Acq On : 18 Sep 2020 1:34
 Operator : JC/MD
 Sample : L4021-07
 Misc : 9.68g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B16-5-7

Quant Time: Sep 18 05:56:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	231291	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	428401	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	428394	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	188662	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	187780	47.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.38%	
35) Dibromofluoromethane	7.55	113	136374	45.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.96%	
50) Toluene-d8	10.06	98	549070	49.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.48%	
62) 4-Bromofluorobenzene	12.38	95	213249	52.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.32%	
Target Compounds						
44) Trichloroethene	8.80	130	7374	2.334	ug/l	93
64) Tetrachloroethene	10.60	164	411003	140.897	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN091720\
Data File : VN063443.D
Acq On : 18 Sep 2020 1:34
Operator : JC/MD
Sample : L4021-07
Misc : 9.68g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B16-5-7

Quant Time: Sep 18 05:56:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

