

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091720\
 Data File : VN063446.D
 Acq On : 18 Sep 2020 2:45
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 18 05:46:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	0.566	0.541	4.4	101	0.00
3 P	Chloromethane	0.749	0.643	14.2	94	0.00
4 C	Vinyl Chloride	0.664	0.642	3.3#	100	0.00
5 T	Bromomethane	0.394	0.391	0.8	108	0.00
6 T	Chloroethane	0.408	0.382	6.4	103	0.00
7 T	Trichlorofluoromethane	0.991	0.990	0.1	104	0.00
8 T	Diethyl Ether	0.386	0.362	6.2	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.497	10.8	94	0.00
10 T	Methyl Iodide	0.725	0.720	0.7	102	0.00
11 T	Tert butyl alcohol	0.118	0.101	14.4	90	0.00
12 CM	1,1-Dichloroethene	0.564	0.535	5.1#	99	0.00
13 T	Acrolein	0.044	0.043	2.3	98	0.00
14 T	Allyl chloride	1.116	0.984	11.8	90	0.00
15 T	Acrylonitrile	0.300	0.278	7.3	96	0.00
16 T	Acetone	0.274	0.253	7.7	101	0.00
17 T	Carbon Disulfide	1.641	1.425	13.2	91	0.00
18 T	Methyl Acetate	0.694	0.646	6.9	99	0.00
19 T	Methyl tert-butyl Ether	2.103	1.952	7.2	98	0.00
20 T	Methylene Chloride	0.693	0.624	10.0	99	0.00
21 T	trans-1,2-Dichloroethene	0.625	0.566	9.4	96	0.00
22 T	Diisopropyl ether	2.276	2.098	7.8	98	0.00
23 T	Vinyl Acetate	1.869	1.750	6.4	95	0.00
24 P	1,1-Dichloroethane	1.243	1.148	7.6	96	0.00
25 T	2-Butanone	0.421	0.379	10.0	94	0.00
26 T	2,2-Dichloropropane	1.094	0.740	32.4#	72	0.00
27 T	cis-1,2-Dichloroethene	0.708	0.678	4.2	98	0.00
28 T	Bromochloromethane	0.574	0.524	8.7	99	0.00
29 T	Tetrahydrofuran	0.287	0.257	10.5	96	0.00
30 C	Chloroform	1.247	1.189	4.7#	100	0.00
31 T	Cyclohexane	1.176	0.978	16.8	92	0.00
32 T	1,1,1-Trichloroethane	1.109	1.055	4.9	100	0.00
33 S	1,2-Dichloroethane-d4	0.851	0.802	5.8	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	107	0.00
35 S	Dibromofluoromethane	0.346	0.317	8.4	102	0.00
36 T	1,1-Dichloropropene	0.533	0.482	9.6	97	0.00
37 T	Ethyl Acetate	0.523	0.471	9.9	94	0.00
38 T	Carbon Tetrachloride	0.544	0.507	6.8	98	0.00
39 T	Methylcyclohexane	0.578	0.499	13.7	92	0.00
40 TM	Benzene	1.521	1.415	7.0	99	0.00
41 T	Methacrylonitrile	0.235	0.217	7.7	96	-0.01
42 TM	1,2-Dichloroethane	0.612	0.575	6.0	100	0.00
43 T	Isopropyl Acetate	0.935	0.834	10.8	95	0.00
44 TM	Trichloroethene	0.369	0.352	4.6	103	0.00
45 C	1,2-Dichloropropane	0.414	0.379	8.5#	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091720\
 Data File : VN063446.D
 Acq On : 18 Sep 2020 2:45
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 18 05:46:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 16 13:05:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.272	0.253	7.0	98	0.00
47 T	Bromodichloromethane	0.565	0.530	6.2	99	0.00
48 T	Methyl methacrylate	0.459	0.409	10.9	94	0.00
49 T	1,4-Dioxane	0.007	0.006	14.3	93	0.00
50 S	Toluene-d8	1.302	1.218	6.5	103	0.00
51 T	4-Methyl-2-Pentanone	0.523	0.484	7.5	98	0.00
52 CM	Toluene	0.914	0.889	2.7#	101	0.00
53 T	t-1,3-Dichloropropene	0.631	0.569	9.8	95	0.00
54 T	cis-1,3-Dichloropropene	0.666	0.596	10.5	92	0.00
55 T	1,1,2-Trichloroethane	0.363	0.350	3.6	101	0.00
56 T	Ethyl methacrylate	0.567	0.543	4.2	100	0.00
57 T	1,3-Dichloropropane	0.655	0.611	6.7	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.297	0.280	5.7	103	0.00
59 T	2-Hexanone	0.379	0.352	7.1	97	0.00
60 T	Dibromochloromethane	0.410	0.392	4.4	100	0.00
61 T	1,2-Dibromoethane	0.377	0.369	2.1	103	0.00
62 S	4-Bromofluorobenzene	0.477	0.472	1.0	108	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	111	0.00
64 T	Tetrachloroethene	0.340	0.326	4.1	108	0.00
65 PM	Chlorobenzene	1.061	0.988	6.9	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.389	0.368	5.4	102	0.00
67 C	Ethyl Benzene	1.999	1.873	6.3#	103	0.00
68 T	m/p-Xylenes	0.721	0.690	4.3	104	0.00
69 T	o-Xylene	0.686	0.659	3.9	104	0.00
70 T	Styrene	1.151	1.151	0.0	104	0.00
71 P	Bromoform	0.290	0.262	9.7	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
73 T	Isopropylbenzene	4.007	3.702	7.6	102	0.00
74 T	N-amyl acetate	1.897	1.998	-5.3	116	0.00
75 P	1,1,2,2-Tetrachloroethane	1.241	1.109	10.6	102	0.00
76 T	1,2,3-Trichloropropane	1.238	1.029	16.9	101	0.00
77 T	Bromobenzene	0.901	1.138	-26.3#	140	0.00
78 T	n-propylbenzene	4.667	4.327	7.3	103	0.00
79 T	2-Chlorotoluene	2.797	2.644	5.5	106	0.00
80 T	1,3,5-Trimethylbenzene	3.341	3.212	3.9	105	0.00
81 T	trans-1,4-Dichloro-2-butene	0.451	0.350	22.4#	88	0.00
82 T	4-Chlorotoluene	2.956	2.727	7.7	104	0.00
83 T	tert-Butylbenzene	2.758	2.590	6.1	104	0.00
84 T	1,2,4-Trimethylbenzene	3.375	3.181	5.7	102	0.00
85 T	sec-Butylbenzene	3.691	3.403	7.8	102	0.00
86 T	p-Isopropyltoluene	3.256	3.127	4.0	105	0.00
87 T	1,3-Dichlorobenzene	1.671	1.579	5.5	105	0.00
88 T	1,4-Dichlorobenzene	1.698	1.565	7.8	102	0.00
89 T	n-Butylbenzene	3.001	2.804	6.6	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091720\
Data File : VN063446.D
Acq On : 18 Sep 2020 2:45
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Sep 18 05:46:14 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 16 13:05:20 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.620	0.542	12.6	97	0.00
91 T	1,2-Dichlorobenzene	1.640	1.556	5.1	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.295	0.256	13.2	95	0.00
93 T	1,2,4-Trichlorobenzene	0.967	0.928	4.0	102	0.00
94 T	Hexachlorobutadiene	0.400	0.390	2.5	106	0.00
95 T	Naphthalene	3.194	3.173	0.7	105	0.00
96 T	1,2,3-Trichlorobenzene	0.934	0.877	6.1	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6