


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN091720\
 Data File : VN063433.D
 Acq On : 17 Sep 2020 20:48
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091620W.M
 Title : SW846 8260
 Last Update : Wed Sep 16 13:05:20 2020



AutoFind: Scans 3307, 3308, 3309; Background Corrected with Scan 3294

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.0	11983	PASS
75	95	30	60	53.2	26578	PASS
95	95	100	100	100.0	49989	PASS
96	95	5	9	6.6	3282	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	72.4	36200	PASS
175	174	5	9	8.4	3049	PASS
176	174	95	101	96.9	35090	PASS
177	176	5	9	5.4	1885	PASS