

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
 Data File : VN068517.D
 Acq On : 17 Sep 2021 14:03
 Operator : JC/MD
 Sample : M3781-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW180D-20210915

Manual Integrations
 APPROVED

MMDadoda
 9/21/2021 2:08:01 PM

Quant Time: Sep 20 03:56:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	514221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	764886	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	779588	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	295843	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	257966	54.689	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.380%	
35) Dibromofluoromethane	8.024	113	236445	51.253	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.500%	
50) Toluene-d8	10.446	98	967643	51.061	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.120%	
62) 4-Bromofluorobenzene	12.733	95	284310	45.953	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 91.900%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.229	101	2025m	0.459	ug/l	
44) Trichloroethene	9.220	130	14537	2.432	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
Data File : VN068517.D
Acq On : 17 Sep 2021 14:03
Operator : JC/MD
Sample : M3781-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW180D-20210915

Manual Integrations
APPROVED

MMDadoda
9/21/2021 2:08:01 PM

Quant Time: Sep 20 03:56:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

