

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
 Data File : VN068519.D
 Acq On : 17 Sep 2021 14:53
 Operator : JC/MD
 Sample : M3781-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE135D3-20210915

Manual Integrations
 APPROVED

MMDadoda
 9/21/2021 2:08:04 PM

Quant Time: Sep 20 03:57:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	484722	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	706099	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	640689	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	237682	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	241434	54.299	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 108.600%			
35) Dibromofluoromethane	8.024	113	223020	52.368	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 104.740%			
50) Toluene-d8	10.446	98	836981	48.068	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 96.140%			
62) 4-Bromofluorobenzene	12.733	95	243099	42.869	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 85.740%			
Target Compounds					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.218	101	29599	7.116	ug/l #	67
12) 1,1-Dichloroethene	4.189	96	1981m	0.520	ug/l	
27) cis-1,2-Dichloroethene	7.329	96	18771	3.676	ug/l	79
44) Trichloroethene	9.220	130	100574	18.229	ug/l	96
45) 1,2-Dichloropropane	9.494	63	20282	4.875	ug/l	95
64) Tetrachloroethene	10.982	164	127109	22.384	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
Data File : VN068519.D
Acq On : 17 Sep 2021 14:53
Operator : JC/MD
Sample : M3781-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE135D3-20210915

Manual Integrations
APPROVED

MMDadoda
9/21/2021 2:08:04 PM

Quant Time: Sep 20 03:57:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

