

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
 Data File : VN068522.D
 Acq On : 17 Sep 2021 16:09
 Operator : JC/MD
 Sample : M3781-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW180D1-20210915

Manual Integrations
 APPROVED

MMDadoda
 9/21/2021 2:08:08 PM

Quant Time: Sep 20 03:58:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	473434	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	718380	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	625336	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	230812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	241105	55.518	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 111.040%	
35) Dibromofluoromethane	8.024	113	224288	51.766	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.540%	
50) Toluene-d8	10.446	98	846836	47.822	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 95.640%	
62) 4-Bromofluorobenzene	12.734	95	237567	41.342	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 82.680%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.075	85	2454	0.635	ug/l	95
9) 1,1,2-Trichlorotrifluo...	4.226	101	8073m	1.987	ug/l	
12) 1,1-Dichloroethene	4.189	96	4175	1.122	ug/l	95
27) cis-1,2-Dichloroethene	7.332	96	4404	0.883	ug/l	84
30) Chloroform	7.817	83	5452	0.665	ug/l #	79
44) Trichloroethene	9.217	130	641494	114.285	ug/l	96
55) 1,1,2-Trichloroethane	10.899	97	2790	0.598	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
Data File : VN068522.D
Acq On : 17 Sep 2021 16:09
Operator : JC/MD
Sample : M3781-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW180D1-20210915

Manual Integrations
APPROVED

MMDadoda
9/21/2021 2:08:08 PM

Quant Time: Sep 20 03:58:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

