

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
 Data File : VN068523.D
 Acq On : 17 Sep 2021 16:35
 Operator : JC/MD
 Sample : M3781-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW179D-20210915

Quant Time: Sep 20 03:58:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	471670	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	753907	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	670573	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	262604	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	240292	55.538	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.080%
35) Dibromofluoromethane	8.024	113	220878	48.576	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.160%
50) Toluene-d8	10.446	98	865482	46.665	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	93.340%
62) 4-Bromofluorobenzene	12.736	95	251151	41.615	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	83.240%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	7.329	96	86019	17.311	ug/l	91
44) Trichloroethene	9.220	130	422864	71.785	ug/l	98
64) Tetrachloroethene	10.979	164	82534	13.887	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
Data File : VN068523.D
Acq On : 17 Sep 2021 16:35
Operator : JC/MD
Sample : M3781-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW179D-20210915

Quant Time: Sep 20 03:58:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

