

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
 Data File : VN068525.D
 Acq On : 17 Sep 2021 17:25
 Operator : JC/MD
 Sample : M3781-09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB01-20210915

Quant Time: Sep 20 03:59:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	469290	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	687129	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	666303	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	260862	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	235362	54.674	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.340%	
35) Dibromofluoromethane	8.027	113	219174	52.886	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.780%	
50) Toluene-d8	10.446	98	877829	51.528	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.060%	
62) 4-Bromofluorobenzene	12.736	95	251269	45.275	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 90.560%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	5.371	59	12875	9.127	ug/l	# 95
16) Acetone	4.296	43	25068	12.025	ug/l	93
44) Trichloroethene	9.217	130	4878	0.909	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
Data File : VN068525.D
Acq On : 17 Sep 2021 17:25
Operator : JC/MD
Sample : M3781-09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB01-20210915

Quant Time: Sep 20 03:59:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

