

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
 Data File : VN068526.D
 Acq On : 17 Sep 2021 17:50
 Operator : JC/MD
 Sample : M3781-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB01-20210915

Quant Time: Sep 20 03:59:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	709235	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	931595	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	807949	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	313543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	298008	45.806	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.620%
35) Dibromofluoromethane	8.024	113	291221	51.830	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.660%
50) Toluene-d8	10.443	98	1038859	45.432	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	90.860%
62) 4-Bromofluorobenzene	12.734	95	300963	40.483	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	80.960%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.220	130	2015	0.277	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
Data File : VN068526.D
Acq On : 17 Sep 2021 17:50
Operator : JC/MD
Sample : M3781-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB01-20210915

Quant Time: Sep 20 03:59:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

