

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
 Data File : VN068528.D
 Acq On : 17 Sep 2021 18:41
 Operator : JC/MD
 Sample : M3781-12
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE134D4-20210916

Manual Integrations
 APPROVED

MMDadoda
 9/21/2021 2:08:12 PM

Quant Time: Sep 20 03:59:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	517367	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	754885	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	643802	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	238405	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	241859	50.963	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.920%			
35) Dibromofluoromethane	8.027	113	231250	50.791	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.580%			
50) Toluene-d8	10.446	98	874534	47.060	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 94.120%			
62) 4-Bromofluorobenzene	12.736	95	237216	39.491	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 78.980%			
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	4.223	101	1159756	261.212	ug/l	99
12) 1,1-Dichloroethene	4.194	96	7966m	1.960	ug/l	
27) cis-1,2-Dichloroethene	7.327	96	5221m	0.958	ug/l	
44) Trichloroethene	9.220	130	153162	25.967	ug/l	98
64) Tetrachloroethene	10.979	164	77975	13.665	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
Data File : VN068528.D
Acq On : 17 Sep 2021 18:41
Operator : JC/MD
Sample : M3781-12
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE134D4-20210916

Manual Integrations
APPROVED

MMDadoda
9/21/2021 2:08:12 PM

Quant Time: Sep 20 03:59:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

