

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
 Data File : VN068539.D
 Acq On : 18 Sep 2021 00:34
 Operator : JC/MD
 Sample : M3750-03ME
 Misc : 6.13g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NAPL-AREA-04-(5-7)ME

Quant Time: Sep 20 05:58:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	579909	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.965	114	770530	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	725413	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	331310	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	238200	44.779	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	89.560%		
35) Dibromofluoromethane	8.018	113	233387	50.220	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	100.440%		
50) Toluene-d8	10.443	98	897959	47.318	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	94.640%		
62) 4-Bromofluorobenzene	12.736	95	293830	47.036	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	94.080%		
Target Compounds						Qvalue
4) Vinyl Chloride	2.448	62	9215	1.283	ug/l	94
27) cis-1,2-Dichloroethene	7.316	96	132818	21.740	ug/l	89
44) Trichloroethene	9.212	130	12129	2.015	ug/l	96
64) Tetrachloroethene	10.979	164	10210	1.588	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
Data File : VN068539.D
Acq On : 18 Sep 2021 00:34
Operator : JC/MD
Sample : M3750-03ME
Misc : 6.13g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NAPL-AREA-04-(5-7)ME

Quant Time: Sep 20 05:58:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

