

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
 Data File : VN068541.D
 Acq On : 18 Sep 2021 1:25
 Operator : JC/MD
 Sample : M3750-06ME
 Misc : 5.92g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NAPL-AREA-04-(8-10)-END-POINTM

Quant Time: Sep 20 05:59:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	587382	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.965	114	802590	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	746967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	338114	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	250238	46.443	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.880%
35) Dibromofluoromethane	8.018	113	233764	48.292	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.580%
50) Toluene-d8	10.443	98	922234	46.706	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	93.420%
62) 4-Bromofluorobenzene	12.736	95	299581	46.129	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.260%
Target Compounds						
27) cis-1,2-Dichloroethene	7.319	96	10557	1.706	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
Data File : VN068541.D
Acq On : 18 Sep 2021 1:25
Operator : JC/MD
Sample : M3750-06ME
Misc : 5.92g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NAPL-AREA-04-(8-10)-END-POINTM

Quant Time: Sep 20 05:59:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

