

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
 Data File : VN068549.D
 Acq On : 18 Sep 2021 4:47
 Operator : JC/MD
 Sample : M3798-02
 Misc : 6.35g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NAPL-06-(30-32)

Quant Time: Sep 20 06:01:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	620280	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	826121	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	860311	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	443048	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	256571	45.093	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.180%
35) Dibromofluoromethane	8.021	113	245428	49.257	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.520%
50) Toluene-d8	10.443	98	956571	47.037	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.080%
62) 4-Bromofluorobenzene	12.733	95	367022	54.113	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.220%
Target Compounds						
					Qvalue	
64) Tetrachloroethene	10.979	164	24091	3.159	ug/l	95
80) 1,3,5-Trimethylbenzene	13.061	105	63580	2.483	ug/l	99
84) 1,2,4-Trimethylbenzene	13.369	105	137148	5.585	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
Data File : VN068549.D
Acq On : 18 Sep 2021 4:47
Operator : JC/MD
Sample : M3798-02
Misc : 6.35g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NAPL-06-(30-32)

Quant Time: Sep 20 06:01:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

