

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
 Data File : VN068550.D
 Acq On : 18 Sep 2021 5:12
 Operator : JC/MD
 Sample : M3798-03
 Misc : 6.00g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NAPL-07-(15-18)

Quant Time: Sep 20 06:02:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 17 11:02:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	590102	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	811249	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	763486	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	368995	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	255406	47.184	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.360%		
35) Dibromofluoromethane	8.018	113	242090	49.478	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	98.960%		
50) Toluene-d8	10.443	98	927917	46.508	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	93.020%		
62) 4-Bromofluorobenzene	12.733	95	328943	49.751	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	99.500%		
Target Compounds						
64) Tetrachloroethene	10.977	164	27795	4.108	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
Data File : VN068550.D
Acq On : 18 Sep 2021 5:12
Operator : JC/MD
Sample : M3798-03
Misc : 6.00g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NAPL-07-(15-18)

Quant Time: Sep 20 06:02:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
Quant Title : SW846 8260
QLast Update : Fri Sep 17 11:02:11 2021
Response via : Initial Calibration

