


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091721\
 Data File : VN068510.D
 Acq On : 17 Sep 2021 9:16
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N090721W.M
 Title : SW846 8260
 Last Update : Fri Sep 17 11:02:11 2021



AutoFind: Scans 4005, 4006, 4007; Background Corrected with Scan 3989

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	13700	PASS
75	95	30	60	50.6	41531	PASS
95	95	100	100	100.0	82005	PASS
96	95	5	9	6.5	5349	PASS
173	174	0.00	2	1.3	922	PASS
174	95	50	100	89.9	73725	PASS
175	174	5	9	7.5	5550	PASS
176	174	95	101	95.9	70667	PASS
177	176	5	9	6.6	4655	PASS