

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091722\
 Data File : VN074558.D
 Acq On : 17 Sep 2022 16:04
 Operator : JC\MD
 Sample : N4549-24
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB090822

Quant Time: Sep 19 02:56:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	257431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	500532	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	510876	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	218469	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	209781	48.028	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.060%
35) Dibromofluoromethane	7.951	113	173741	52.254	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.500%
50) Toluene-d8	10.381	98	619128	48.148	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.300%
62) 4-Bromofluorobenzene	12.674	95	236787	53.765	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.540%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	5932	2.122	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091722\
Data File : VN074558.D
Acq On : 17 Sep 2022 16:04
Operator : JC\MD
Sample : N4549-24
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB090822

Quant Time: Sep 19 02:56:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

