

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091722\
 Data File : VN074559.D
 Acq On : 17 Sep 2022 16:29
 Operator : JC\MD
 Sample : N4549-23
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB090822

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/19/2022
 Supervised By : Mahesh Dadoda 09/19/2022

Quant Time: Sep 19 02:56:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	261412	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.904	114	514043	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	507208	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	221183	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	205769	46.392	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.780%		
35) Dibromofluoromethane	7.951	113	172577	50.540	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.080%		
50) Toluene-d8	10.380	98	638429	48.344	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.680%		
62) 4-Bromofluorobenzene	12.674	95	237983	52.617	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	105.240%		
Target Compounds						
16) Acetone	4.222	43	5588	1.969	ug/l	Qvalue 90
25) 2-Butanone	7.245	43	7864m	1.669	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091722\
Data File : VN074559.D
Acq On : 17 Sep 2022 16:29
Operator : JC\MD
Sample : N4549-23
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB090822

Quant Time: Sep 19 02:56:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 09/19/2022
Supervised By :Mahesh Dadoda 09/19/2022

