

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091722\
 Data File : VN074563.D
 Acq On : 17 Sep 2022 18:06
 Operator : JC\MD
 Sample : N4688-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-GW-463-465

Quant Time: Sep 19 02:57:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	259958	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	514294	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	512193	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	221371	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	213124	48.319	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.640%
35) Dibromofluoromethane	7.957	113	174986	51.220	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.440%
50) Toluene-d8	10.380	98	644589	48.787	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.580%
62) 4-Bromofluorobenzene	12.674	95	241940	53.465	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.940%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	10685	3.786	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091722\
Data File : VN074563.D
Acq On : 17 Sep 2022 18:06
Operator : JC\MD
Sample : N4688-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW5-GW-463-465

Quant Time: Sep 19 02:57:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

