

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091722\
 Data File : VN074567.D
 Acq On : 17 Sep 2022 19:43
 Operator : JC\MD
 Sample : N4663-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HW-9

Quant Time: Sep 19 02:58:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	269120	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	518798	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	521182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	256096	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	211344	46.284	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.560%
35) Dibromofluoromethane	7.951	113	173760	50.420	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.840%
50) Toluene-d8	10.380	98	644767	48.377	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.760%
62) 4-Bromofluorobenzene	12.674	95	249071	54.563	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.120%
Target Compounds						
16) Acetone	4.240	43	10106	3.459	ug/l	Qvalue # 64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091722\
Data File : VN074567.D
Acq On : 17 Sep 2022 19:43
Operator : JC\MD
Sample : N4663-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HW-9

Quant Time: Sep 19 02:58:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

