

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091722\
 Data File : VN074571.D
 Acq On : 17 Sep 2022 21:21
 Operator : JC\MD
 Sample : N4647-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1037-MW-11(15)

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/19/2022
 Supervised By : Mahesh Dadoda 09/19/2022

Quant Time: Sep 19 02:59:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 15 19:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	255910	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	511611	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	520327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	228103	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	207362	47.756	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	95.520%		
35) Dibromofluoromethane	7.957	113	168888	49.694	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.380%		
50) Toluene-d8	10.380	98	620830	47.235	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	94.480%		
62) 4-Bromofluorobenzene	12.674	95	243724	54.142	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	108.280%		
Target Compounds						Qvalue
4) Vinyl Chloride	2.404	62	10821	2.158	ug/l #	64
16) Acetone	4.222	43	40277	14.496	ug/l	98
21) trans-1,2-Dichloroethene	5.493	96	2295m	1.044	ug/l	
27) cis-1,2-Dichloroethene	7.245	96	197954	41.775	ug/l	97
29) Tetrahydrofuran	7.628	42	3774	1.244	ug/l #	73
44) Trichloroethene	9.151	130	566580	140.761	ug/l	86
64) Tetrachloroethene	10.916	164	1471181	411.996	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091722\
Data File : VN074571.D
Acq On : 17 Sep 2022 21:21
Operator : JC\MD
Sample : N4647-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1037-MW-11(15)

Manual Integrations
APPROVED

Reviewed By : John Carlone 09/19/2022
Supervised By : Mahesh Dadoda 09/19/2022

Quant Time: Sep 19 02:59:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091522W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 15 19:20:33 2022
Response via : Initial Calibration

