

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
 Data File : VN083956.D
 Acq On : 17 Sep 2024 19:30
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Sep 18 01:54:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	124982	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	223386	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	199716	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	99485	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	110870	58.831	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 117.660%	
35) Dibromofluoromethane	8.165	113	77654	54.300	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.600%	
50) Toluene-d8	10.565	98	280498	56.736	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 113.480%#	
62) 4-Bromofluorobenzene	12.847	95	108765	53.323	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 106.640%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	67147	46.261	ug/l	94
3) Chloromethane	2.359	50	73796	49.780	ug/l	99
4) Vinyl Chloride	2.512	62	75359	47.367	ug/l	97
5) Bromomethane	2.942	94	47138	51.604	ug/l	97
6) Chloroethane	3.112	64	50613	45.352	ug/l #	89
7) Trichlorofluoromethane	3.495	101	130986	48.504	ug/l	100
8) Diethyl Ether	3.965	74	45160	49.887	ug/l	94
9) 1,1,2-Trichlorotrifluo...	4.371	101	70017	48.915	ug/l	98
10) Methyl Iodide	4.589	142	88618	48.679	ug/l	95
11) Tert butyl alcohol	5.530	59	85260	337.852	ug/l	98
12) 1,1-Dichloroethene	4.336	96	67251	49.749	ug/l	98
13) Acrolein	4.177	56	62685	251.753	ug/l	99
14) Allyl chloride	5.018	41	112155	46.562	ug/l	93
15) Acrylonitrile	5.718	53	194812	283.861	ug/l	99
16) Acetone	4.430	43	176405	233.503	ug/l	99
17) Carbon Disulfide	4.712	76	168675	44.606	ug/l	99
18) Methyl Acetate	5.018	43	96069	61.458	ug/l	98
19) Methyl tert-butyl Ether	5.794	73	257840	52.710	ug/l	100
20) Methylene Chloride	5.277	84	78863	51.012	ug/l	92
21) trans-1,2-Dichloroethene	5.789	96	70568	48.710	ug/l	92
22) Diisopropyl ether	6.671	45	269603	54.372	ug/l	99
23) Vinyl Acetate	6.600	43	1025740	282.482	ug/l	98
24) 1,1-Dichloroethane	6.571	63	146434	51.260	ug/l	99
25) 2-Butanone	7.482	43	273797	275.742	ug/l	100
26) 2,2-Dichloropropane	7.488	77	122185	46.444	ug/l	99
27) cis-1,2-Dichloroethene	7.482	96	90248	52.477	ug/l	100
28) Bromochloromethane	7.812	49	71192	59.178	ug/l	94
29) Tetrahydrofuran	7.835	42	175011	298.083	ug/l	96
30) Chloroform	7.965	83	157064	51.789	ug/l	98
31) Cyclohexane	8.259	56	118868	43.642	ug/l	97
32) 1,1,1-Trichloroethane	8.165	97	143724	51.539	ug/l	96
36) 1,1-Dichloropropene	8.371	75	105469	48.999	ug/l	99
37) Ethyl Acetate	7.559	43	108399	55.481	ug/l	100
38) Carbon Tetrachloride	8.359	117	123190	50.447	ug/l	99
39) Methylcyclohexane	9.600	83	111097	45.345	ug/l	99
40) Benzene	8.606	78	320770	49.511	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
 Data File : VN083956.D
 Acq On : 17 Sep 2024 19:30
 Operator : JC\MD
 Sample : VSTDC0050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDC0050EC

Quant Time: Sep 18 01:54:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	63317	54.792	ug/l	99
42) 1,2-Dichloroethane	8.671	62	125755	51.946	ug/l	100
43) Isopropyl Acetate	8.688	43	209016	59.034	ug/l	99
44) Trichloroethene	9.353	130	72745	48.487	ug/l	91
45) 1,2-Dichloropropane	9.618	63	81829	52.258	ug/l	99
46) Dibromomethane	9.706	93	59040	53.830	ug/l	98
47) Bromodichloromethane	9.888	83	128081	53.959	ug/l	97
48) Methyl methacrylate	9.682	41	95362	56.770	ug/l	96
49) 1,4-Dioxane	9.694	88	32684	1182.761	ug/l #	98
51) 4-Methyl-2-Pentanone	10.447	43	579926	300.898	ug/l	97
52) Toluene	10.629	92	203602	51.378	ug/l	97
53) t-1,3-Dichloropropene	10.835	75	126906	52.746	ug/l	100
54) cis-1,3-Dichloropropene	10.312	75	132569	52.454	ug/l	99
55) 1,1,2-Trichloroethane	11.018	97	77795	52.584	ug/l	97
56) Ethyl methacrylate	10.871	69	131858	57.929	ug/l	97
57) 1,3-Dichloropropane	11.165	76	136392	53.641	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	273177	333.076	ug/l	97
59) 2-Hexanone	11.194	43	424938	298.059	ug/l	97
60) Dibromochloromethane	11.359	129	93822	56.666	ug/l	100
61) 1,2-Dibromoethane	11.465	107	79357	53.784	ug/l	98
64) Tetrachloroethene	11.100	164	64008	48.193	ug/l	95
65) Chlorobenzene	11.888	112	216980	48.883	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.959	131	78376	50.933	ug/l	99
67) Ethyl Benzene	11.965	91	399288	50.638	ug/l	99
68) m/p-Xylenes	12.070	106	295234	100.849	ug/l	99
69) o-Xylene	12.400	106	144653	51.042	ug/l	99
70) Styrene	12.412	104	250352	53.827	ug/l	99
71) Bromoform	12.576	173	59898	57.764	ug/l #	99
73) Isopropylbenzene	12.694	105	377533	49.051	ug/l	100
74) N-amyl acetate	12.494	43	180019	54.819	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.935	83	115647	51.138	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	150091	74.769	ug/l	61
77) Bromobenzene	12.976	156	86561	46.721	ug/l	98
78) n-propylbenzene	13.035	91	447993	50.025	ug/l	98
79) 2-Chlorotoluene	13.123	91	277433	48.577	ug/l	99
80) 1,3,5-Trimethylbenzene	13.170	105	319911	49.300	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.735	75	39154	51.084	ug/l	95
82) 4-Chlorotoluene	13.223	91	279952	48.846	ug/l	99
83) tert-Butylbenzene	13.435	119	277668	48.908	ug/l	100
84) 1,2,4-Trimethylbenzene	13.482	105	323422	49.617	ug/l	99
85) sec-Butylbenzene	13.617	105	378644	50.289	ug/l	99
86) p-Isopropyltoluene	13.729	119	314842	49.883	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	166200	47.533	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	164485	46.459	ug/l	100
89) n-Butylbenzene	14.053	91	274759	48.486	ug/l	98
90) Hexachloroethane	14.329	117	60035	47.318	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	162035	48.018	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.717	75	25065	55.600	ug/l	98
93) 1,2,4-Trichlorobenzene	15.394	180	80371	44.384	ug/l	98
94) Hexachlorobutadiene	15.500	225	34464	44.816	ug/l	99
95) Naphthalene	15.641	128	269153	48.856	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	78714	45.484	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
Data File : VN083956.D
Acq On : 17 Sep 2024 19:30
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC

Quant Time: Sep 18 01:54:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:168 Resp: 124982
Ion Ratio Lower Upper
168 100
99 70.6 54.2 81.2



#2
Dichlorodifluoromethane
Concen: 46.261 ug/l
RT: 2.124 min Scan# 29
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 85 Resp: 67147
Ion Ratio Lower Upper
85 100
87 34.5 15.7 47.0









#3
Chloromethane
Concen: 49.780 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 50 Resp: 73796
Ion Ratio Lower Upper
50 100
52 34.4 27.2 40.8



#4
Vinyl Chloride
Concen: 47.367 ug/l
RT: 2.512 min Scan# 95
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 62 Resp: 75359
Ion Ratio Lower Upper
62 100
64 32.5 24.7 37.1









#5
Bromomethane
Concen: 51.604 ug/l
RT: 2.942 min Scan# 11
Delta R.T. 0.035 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDC050EC


Tgt Ion: 94 Resp: 47138
Ion Ratio Lower Upper
94 100
96 91.8 76.0 114.0



#6
Chloroethane
Concen: 45.352 ug/l
RT: 3.112 min Scan# 197
Delta R.T. 0.029 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 64 Resp: 50613
Ion Ratio Lower Upper
64 100
66 36.0 23.9 35.9#









#7
Trichlorofluoromethane
Concen: 48.504 ug/l
RT: 3.495 min Scan# 20
Delta R.T. 0.023 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 130986
Ion Ratio Lower Upper
101 100
103 64.2 51.4 77.0



#8
Diethyl Ether
Concen: 49.887 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.012 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 74 Resp: 45160
Ion Ratio Lower Upper
74 100
45 99.4 46.7 140.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.915 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.018 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:101 Resp: 70017
Ion Ratio Lower Upper
101 100
85 48.0 36.2 54.4
151 72.2 57.7 86.5



#10
Methyl Iodide
Concen: 48.679 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.012 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion:142 Resp: 88618
Ion Ratio Lower Upper
142 100
127 50.6 37.5 56.3
141 14.8 11.0 16.4









#11
Tert butyl alcohol
Concen: 337.852 ug/l
RT: 5.530 min Scan# 608
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 59 Resp: 85260
Ion Ratio Lower Upper
59 100
57 10.2 8.7 13.1



#12
1,1-Dichloroethene
Concen: 49.749 ug/l
RT: 4.336 min Scan# 405
Delta R.T. 0.006 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 96 Resp: 67251
Ion Ratio Lower Upper
96 100
61 176.8 143.4 215.0
98 63.3 52.7 79.1









#13
Acrolein
Concen: 251.753 ug/l
RT: 4.177 min Scan# 31
Delta R.T. 0.006 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 62685
Ion Ratio Lower Upper
56 100
55 72.1 58.2 87.2



#14
Allyl chloride
Concen: 46.562 ug/l
RT: 5.018 min Scan# 521
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 41 Resp: 112155
Ion Ratio Lower Upper
41 100
39 81.0 59.3 88.9
76 37.3 27.7 41.5









#15
Acrylonitrile
Concen: 283.861 ug/l
RT: 5.718 min Scan# 640
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 53 Resp: 194812
Ion Ratio Lower Upper
53 100
52 81.9 66.1 99.1
51 38.0 29.6 44.4



#16
Acetone
Concen: 233.503 ug/l
RT: 4.430 min Scan# 421
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 43 Resp: 176405
Ion Ratio Lower Upper
43 100
58 30.1 23.7 35.5





#17

Carbon Disulfide

Concen: 44.606 ug/l

RT: 4.712 min Scan# 4

Delta R.T. 0.012 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 76 Resp: 168675

Ion Ratio Lower Upper

76 100

78 8.4 7.1 10.7



#18

Methyl Acetate

Concen: 61.458 ug/l

RT: 5.018 min Scan# 521

Delta R.T. -0.006 min

Lab File: VN083956.D




Acq: 17 Sep 2024 19:30


Tgt Ion: 43 Resp: 96069

Ion Ratio Lower Upper

43 100

74 22.9 19.0 28.4





#19

Methyl tert-butyl Ether

Concen: 52.710 ug/l

RT: 5.794 min Scan# 61

Delta R.T. 0.006 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 73 Resp: 257840

Ion Ratio Lower Upper

73 100

57 21.1 17.0 25.6



#20

Methylene Chloride

Concen: 51.012 ug/l

RT: 5.277 min Scan# 565

Delta R.T. 0.006 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30

Tgt Ion: 84 Resp: 78863




Ion Ratio Lower Upper


84 100

49 125.4 93.4 140.0

51 39.5 29.2 43.8

86 65.8 46.9 70.3









#21
trans-1,2-Dichloroethene
Concen: 48.710 ug/l
RT: 5.789 min Scan# 61
Delta R.T. 0.012 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 96 Resp: 70568
Ion Ratio Lower Upper
96 100
61 142.5 106.0 159.0
98 60.0 52.2 78.2



#22
Diisopropyl ether
Concen: 54.372 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 45 Resp: 269603
Ion Ratio Lower Upper
45 100
43 53.4 43.4 65.0
87 27.6 22.6 33.8
59 11.6 9.1 13.7









#23
 Vinyl Acetate
 Concen: 282.482 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. -0.000 min
 Lab File: VN083956.D
 Acq: 17 Sep 2024 19:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 43 Resp: 1025740
 Ion Ratio Lower Upper
 43 100
 86 10.0 8.7 13.1



#24
 1,1-Dichloroethane
 Concen: 51.260 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. 0.012 min
 Lab File: VN083956.D
 Acq: 17 Sep 2024 19:30

Tgt Ion: 63 Resp: 146434
 Ion Ratio Lower Upper
 63 100
 98 6.2 3.5 10.4
 100 4.8 2.5 7.5









#25
2-Butanone
Concen: 275.742 ug/l
RT: 7.482 min Scan# 940
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 273797
Ion Ratio Lower Upper
43 100
72 26.1 20.8 31.2



#26
2,2-Dichloropropane
Concen: 46.444 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 77 Resp: 122185
Ion Ratio Lower Upper
77 100
97 20.8 10.2 30.4









#27
 cis-1,2-Dichloroethene
 Concen: 52.477 ug/l
 RT: 7.482 min Scan# 940
 Delta R.T. -0.000 min
 Lab File: VN083956.D
 Acq: 17 Sep 2024 19:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 96 Resp: 90248
 Ion Ratio Lower Upper
 96 100
 61 153.4 0.0 306.2
 98 63.5 0.0 128.6



#28
 Bromochloromethane
 Concen: 59.178 ug/l
 RT: 7.812 min Scan# 996
 Delta R.T. -0.000 min
 Lab File: VN083956.D
 Acq: 17 Sep 2024 19:30

Tgt Ion: 49 Resp: 71192
 Ion Ratio Lower Upper
 49 100
 129 1.7 0.0 3.4
 130 70.2 60.2 90.2









#29
Tetrahydrofuran
Concen: 298.083 ug/l
RT: 7.835 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	42	Resp:	175011
Ion	Ratio	Lower	Upper
42	100		
72	42.6	35.9	53.9
71	39.8	33.9	50.9



#30
Chloroform
Concen: 51.789 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.006 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion:	83	Resp:	157064
Ion	Ratio	Lower	Upper
83	100		
85	64.5	52.8	79.2









#31
Cyclohexane
Concen: 43.642 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. 0.006 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 56 Resp: 118868
Ion Ratio Lower Upper
56 100
69 29.6 26.9 40.3
84 85.2 69.2 103.8



#32
1,1,1-Trichloroethane
Concen: 51.539 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 97 Resp: 143724
Ion Ratio Lower Upper
97 100
99 62.7 52.3 78.5
61 50.2 38.2 57.2





#33

1,2-Dichloroethane-d4

Concen: 58.831 ug/l

RT: 8.577 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 65 Resp: 110870

Ion Ratio Lower Upper

65 100

67 46.7 0.0 100.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30




Tgt Ion: 114 Resp: 223386


Ion Ratio Lower Upper

114 100

63 21.9 0.0 43.4

88 15.5 0.0 32.6





#35

Dibromofluoromethane

Concen: 54.300 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:113 Resp: 77654

Ion Ratio Lower Upper

113 100

111 103.1 82.6 124.0

192 17.2 14.3 21.5



#36

1,1-Dichloropropene

Concen: 48.999 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.000 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30




Tgt Ion: 75 Resp: 105469


Ion Ratio Lower Upper

75 100

110 32.3 16.7 50.1

77 31.0 25.2 37.8









#37
Ethyl Acetate
Concen: 55.481 ug/l
RT: 7.559 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 108399
Ion Ratio Lower Upper
43 100
61 13.8 11.1 16.7
70 10.9 8.8 13.2



#38
Carbon Tetrachloride
Concen: 50.447 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 117 Resp: 123190
Ion Ratio Lower Upper
117 100
119 94.0 76.1 114.1
121 29.0 24.1 36.1





#39

Methylcyclohexane

Concen: 45.345 ug/l

RT: 9.600 min Scan# 1110

Delta R.T. -0.000 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 83 Resp: 111097

Ion Ratio Lower Upper

83 100

55 78.8 62.2 93.4

98 46.8 37.6 56.4



#40

Benzene

Concen: 49.511 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. 0.006 min

Lab File: VN083956.D




Acq: 17 Sep 2024 19:30


Tgt Ion: 78 Resp: 320770

Ion Ratio Lower Upper

78 100

77 23.7 19.0 28.6









#41
Methacrylonitrile
Concen: 54.792 ug/l
RT: 7.777 min Scan# 990
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	41	Resp:	63317
Ion	Ratio	Lower	Upper
41	100		
39	56.2	45.4	68.2
67	69.1	55.9	83.9
52	30.7	24.1	36.1



#42
1,2-Dichloroethane
Concen: 51.946 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion:	62	Resp:	125755
Ion	Ratio	Lower	Upper
62	100		
98	8.2	0.0	16.6









#43
Isopropyl Acetate
Concen: 59.034 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 209016
Ion Ratio Lower Upper
43 100
61 23.5 19.4 29.0
87 11.7 9.8 14.6



#44
Trichloroethene
Concen: 48.487 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion:130 Resp: 72745
Ion Ratio Lower Upper
130 100
95 109.4 0.0 200.4





#45

1,2-Dichloropropane

Concen: 52.258 ug/l

RT: 9.618 min Scan# 1303

Delta R.T. -0.000 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion: 63 Resp: 81829

Ion Ratio Lower Upper

63 100

65 33.0 25.9 38.9



#46

Dibromomethane

Concen: 53.830 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.000 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30




Tgt Ion: 93 Resp: 59040


Ion Ratio Lower Upper

93 100

95 80.9 65.3 97.9

174 86.1 71.0 106.4









#47
Bromodichloromethane
Concen: 53.959 ug/l
RT: 9.888 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 83 Resp: 128081
Ion Ratio Lower Upper
83 100
85 61.9 51.8 77.8
127 7.6 6.6 9.8



#48
Methyl methacrylate
Concen: 56.770 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. 0.006 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 41 Resp: 95362
Ion Ratio Lower Upper
41 100
69 76.6 64.3 96.5
39 63.4 48.9 73.3





#49

1,4-Dioxane

Concen: 1182.761 ug/l

RT: 9.694 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion: 88 Resp: 32684

Ion Ratio Lower Upper

88 100

43 35.7 0.0 0.0#

58 72.8 57.1 85.7



#50

Toluene-d8

Concen: 56.736 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN083956.D




Acq: 17 Sep 2024 19:30


Tgt Ion: 98 Resp: 280498

Ion Ratio Lower Upper

98 100

100 64.4 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 300.898 ug/l
RT: 10.447 min Scan# 1443
Delta R.T. 0.006 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 579926
Ion Ratio Lower Upper
43 100
58 37.7 31.6 47.4



#52
Toluene
Concen: 51.378 ug/l
RT: 10.629 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 92 Resp: 203602
Ion Ratio Lower Upper
92 100
91 172.9 134.7 202.1









#53
t-1,3-Dichloropropene
Concen: 52.746 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 75 Resp: 126906
Ion Ratio Lower Upper
75 100
77 32.4 25.8 38.6



#54
cis-1,3-Dichloropropene
Concen: 52.454 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 75 Resp: 132569
Ion Ratio Lower Upper
75 100
77 31.2 25.0 37.6
39 54.3 42.6 64.0









#55
1,1,2-Trichloroethane
Concen: 52.584 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	97	Resp:	77795
Ion	Ratio	Lower	Upper
97	100		
83	90.4	67.5	101.3
85	55.3	44.2	66.2
99	61.4	50.1	75.1



#56
Ethyl methacrylate
Concen: 57.929 ug/l
RT: 10.871 min Scan# 1516
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion:	69	Resp:	131858
Ion	Ratio	Lower	Upper
69	100		
41	69.2	53.9	80.9
39	42.9	32.6	48.8









#57
1,3-Dichloropropane
Concen: 53.641 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. 0.006 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 76 Resp: 136392
Ion Ratio Lower Upper
76 100
78 32.1 25.8 38.6



#58
2-Chloroethyl Vinyl ether
Concen: 333.076 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 63 Resp: 273177
Ion Ratio Lower Upper
63 100
106 27.6 23.3 34.9









#59
2-Hexanone
Concen: 298.059 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 43 Resp: 424938
Ion Ratio Lower Upper
43 100
58 52.4 27.2 81.5



#60
Dibromochloromethane
Concen: 56.666 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion:129 Resp: 93822
Ion Ratio Lower Upper
129 100
127 77.6 39.0 116.9









#61
1,2-Dibromoethane
Concen: 53.784 ug/l
RT: 11.465 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:107 Resp: 79357
Ion Ratio Lower Upper
107 100
109 92.8 75.6 113.4



#62
4-Bromofluorobenzene
Concen: 53.323 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 95 Resp: 108765
Ion Ratio Lower Upper
95 100
174 71.9 0.0 139.8
176 68.3 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:117 Resp: 199716
Ion Ratio Lower Upper
117 100
82 58.5 47.4 71.0
119 32.5 26.3 39.5



#64
Tetrachloroethene
Concen: 48.193 ug/l
RT: 11.100 min Scan# 1555
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion:164 Resp: 64008
Ion Ratio Lower Upper
164 100
166 120.7 103.7 155.5
129 99.6 79.9 119.9
131 92.8 78.4 117.6









#65
Chlorobenzene
Concen: 48.883 ug/l
RT: 11.888 min Scan# 112
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:112 Resp: 216980
Ion Ratio Lower Upper
112 100
114 32.5 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 50.933 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion:131 Resp: 78376
Ion Ratio Lower Upper
131 100
133 96.4 47.5 142.6
119 63.9 31.6 94.8









#67
Ethyl Benzene
Concen: 50.638 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 399288
Ion Ratio Lower Upper
91 100
106 29.8 24.3 36.5



#68
m/p-Xylenes
Concen: 100.849 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion:106 Resp: 295234
Ion Ratio Lower Upper
106 100
91 213.0 169.2 253.8





#69

o-Xylene

Concen: 51.042 ug/l

RT: 12.400 min Scan# 1776

Delta R.T. 0.006 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30

Instrument :

MSVOA_N

ClientSampleId :





VSTDCCC050EC

Tgt Ion:106 Resp: 144653

Ion Ratio Lower Upper

106 100

91 226.8 112.9 338.6



#70

Styrene

Concen: 53.827 ug/l

RT: 12.412 min Scan# 1778

Delta R.T. -0.000 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30




Tgt Ion:104 Resp: 250352


Ion Ratio Lower Upper

104 100

78 54.6 43.3 64.9

103 54.6 43.4 65.2









#71
Bromoform
Concen: 57.764 ug/l
RT: 12.576 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:173 Resp: 59898
Ion Ratio Lower Upper
173 100
175 48.7 24.1 72.2
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.006 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion:152 Resp: 99485
Ion Ratio Lower Upper
152 100
115 63.5 32.2 96.6
150 173.6 0.0 348.6









#73
Isopropylbenzene
Concen: 49.051 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:105 Resp: 377533
Ion Ratio Lower Upper
105 100
120 25.3 12.6 37.8



#74
N-ethyl acetate
Concen: 54.819 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 43 Resp: 180019
Ion Ratio Lower Upper
43 100
70 40.9 34.5 51.7
55 24.6 18.6 28.0
61 23.4 19.2 28.8









#75
 1,1,2,2-Tetrachloroethane
 Concen: 51.138 ug/l
 RT: 12.935 min Scan# 1867
 Delta R.T. -0.000 min
 Lab File: VN083956.D
 Acq: 17 Sep 2024 19:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion: 83 Resp: 115647
 Ion Ratio Lower Upper
 83 100
 131 9.8 5.2 15.6
 85 62.9 31.9 95.9



#76
 1,2,3-Trichloropropane
 Concen: 74.769 ug/l
 RT: 12.994 min Scan# 1877
 Delta R.T. -0.000 min
 Lab File: VN083956.D
 Acq: 17 Sep 2024 19:30

Tgt Ion: 75 Resp: 150091
 Ion Ratio Lower Upper
 75 100
 77 133.5 95.2 285.6









#77
Bromobenzene
Concen: 46.721 ug/l
RT: 12.976 min Scan# 1874
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:156 Resp: 86561
Ion Ratio Lower Upper
156 100
77 231.5 114.3 343.1
158 96.3 48.9 146.7



#78
n-propylbenzene
Concen: 50.025 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 91 Resp: 447993
Ion Ratio Lower Upper
91 100
120 21.2 11.0 33.0









#79
2-Chlorotoluene
Concen: 48.577 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 277433
Ion Ratio Lower Upper
91 100
126 31.4 15.4 46.2



#80
1,3,5-Trimethylbenzene
Concen: 49.300 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 105 Resp: 319911
Ion Ratio Lower Upper
105 100
120 46.6 23.4 70.3









#81
trans-1,4-Dichloro-2-butene
Concen: 51.084 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:	75	Resp:	39154
Ion	Ratio	Lower	Upper
75	100		
53	108.7	81.2	121.8
89	44.8	35.7	53.5



#82
4-Chlorotoluene
Concen: 48.846 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion:	91	Resp:	279952
Ion	Ratio	Lower	Upper
91	100		
126	31.6	15.4	46.2





#83

tert-Butylbenzene

Concen: 48.908 ug/l

RT: 13.435 min Scan# 1952

Delta R.T. -0.000 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:119 Resp: 277668

Ion Ratio Lower Upper

119 100

91 69.2 34.5 103.6

134 25.9 12.8 38.4



#84

1,2,4-Trimethylbenzene

Concen: 49.617 ug/l

RT: 13.482 min Scan# 1960

Delta R.T. -0.000 min

Lab File: VN083956.D




Acq: 17 Sep 2024 19:30


Tgt Ion:105 Resp: 323422

Ion Ratio Lower Upper

105 100

120 44.0 21.8 65.4









#85
 sec-Butylbenzene
 Concen: 50.289 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN083956.D
 Acq: 17 Sep 2024 19:30




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDCCC050EC


Tgt Ion:105 Resp: 378644
 Ion Ratio Lower Upper
 105 100
 134 18.9 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 49.883 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN083956.D
 Acq: 17 Sep 2024 19:30

Tgt Ion:119 Resp: 314842
 Ion Ratio Lower Upper
 119 100
 134 25.0 12.8 38.3
 91 26.2 12.9 38.7





#87

1,3-Dichlorobenzene

Concen: 47.533 ug/l

RT: 13.735 min Scan# 2002

Delta R.T. 0.006 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:146 Resp: 166200

Ion Ratio Lower Upper

146 100

111 42.7 21.7 65.1

148 62.2 31.7 95.1



#88

1,4-Dichlorobenzene

Concen: 46.459 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30




Tgt Ion:146 Resp: 164485


Ion Ratio Lower Upper

146 100

111 42.7 21.3 64.0

148 64.3 32.0 96.2









#89
n-Butylbenzene
Concen: 48.486 ug/l
RT: 14.053 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion: 91 Resp: 274759
Ion Ratio Lower Upper
91 100
92 52.0 26.7 80.1
134 23.5 12.2 36.4



#90
Hexachloroethane
Concen: 47.318 ug/l
RT: 14.329 min Scan# 2104
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 117 Resp: 60035
Ion Ratio Lower Upper
117 100
201 67.1 32.8 98.3









#91
1,2-Dichlorobenzene
Concen: 48.018 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30




Instrument :
MSVOA_N
ClientSampleId :
VSTDCCC050EC


Tgt Ion:146 Resp: 162035
Ion Ratio Lower Upper
146 100
111 45.0 22.8 68.3
148 64.0 32.1 96.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 55.600 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. -0.000 min
Lab File: VN083956.D
Acq: 17 Sep 2024 19:30

Tgt Ion: 75 Resp: 25065
Ion Ratio Lower Upper
75 100
155 71.1 35.9 107.7
157 91.9 47.5 142.5





#93

1,2,4-Trichlorobenzene

Concen: 44.384 ug/l

RT: 15.394 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:180 Resp: 80371

Ion Ratio Lower Upper

180 100

182 93.4 48.1 144.3

145 34.2 16.9 50.7



#94

Hexachlorobutadiene

Concen: 44.816 ug/l

RT: 15.500 min Scan# 2303

Delta R.T. -0.000 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30




Tgt Ion:225 Resp: 34464


Ion Ratio Lower Upper

225 100

223 64.4 31.9 95.5

227 64.0 32.0 96.2





#95

Naphthalene

Concen: 48.856 ug/l

RT: 15.641 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30

Instrument :

MSVOA_N

ClientSampleId :

VSTDCCC050EC





Tgt Ion:128 Resp: 269153

Ion Ratio Lower Upper

128 100

127 13.3 10.2 15.2

129 11.3 8.7 13.1



#96

1,2,3-Trichlorobenzene

Concen: 45.484 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. -0.000 min

Lab File: VN083956.D

Acq: 17 Sep 2024 19:30

Tgt Ion:180 Resp: 78714

Ion Ratio Lower Upper

180 100

182 94.8 47.1 141.4

145 35.2 17.8 53.5

