

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
 Data File : VN083936.D
 Acq On : 17 Sep 2024 11:30
 Operator : JC\MD
 Sample : P3970-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT189D1-HYD-20240910

Quant Time: Sep 18 01:45:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	161054	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	271030	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	230805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	95641	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	124183	51.136	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.280%
35) Dibromofluoromethane	8.165	113	87641	50.511	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.020%
50) Toluene-d8	10.565	98	322118	53.701	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.400%
62) 4-Bromofluorobenzene	12.847	95	118503	47.885	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
Data File : VN083936.D
Acq On : 17 Sep 2024 11:30
Operator : JC\MD
Sample : P3970-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT189D1-HYD-20240910

Quant Time: Sep 18 01:45:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

