

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
 Data File : VN083938.D
 Acq On : 17 Sep 2024 12:18
 Operator : JC\MD
 Sample : P3970-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-TB-20240910

Quant Time: Sep 18 02:34:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	158471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	270810	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	227497	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	90002	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	124055	51.916	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.840%
35) Dibromofluoromethane	8.165	113	89285	51.500	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.000%
50) Toluene-d8	10.565	98	323975	54.055	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.100%
62) 4-Bromofluorobenzene	12.847	95	112640	45.552	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
Data File : VN083938.D
Acq On : 17 Sep 2024 12:18
Operator : JC\MD
Sample : P3970-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-TB-20240910

Quant Time: Sep 18 02:34:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

