

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
 Data File : VN083947.D
 Acq On : 17 Sep 2024 15:53
 Operator : JC\MD
 Sample : P3996-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-SOIL-REF-6

Quant Time: Sep 18 01:51:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153777	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	267793	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	225304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	95233	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	118837	51.251	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.500%
35) Dibromofluoromethane	8.171	113	86831	50.649	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.300%
50) Toluene-d8	10.565	98	319029	53.829	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.660%
62) 4-Bromofluorobenzene	12.847	95	114284	46.738	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
Data File : VN083947.D
Acq On : 17 Sep 2024 15:53
Operator : JC\MD
Sample : P3996-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-SOIL-REF-6

Quant Time: Sep 18 01:51:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

