

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
 Data File : VN083950.D
 Acq On : 17 Sep 2024 17:05
 Operator : JC\MD
 Sample : P3996-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Sep 18 01:52:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	151205	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	259755	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	219906	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	92966	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	117505	51.538	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.080%
35) Dibromofluoromethane	8.165	113	83079	49.960	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.920%
50) Toluene-d8	10.565	98	311335	54.157	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.320%
62) 4-Bromofluorobenzene	12.847	95	111635	47.067	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
Data File : VN083950.D
Acq On : 17 Sep 2024 17:05
Operator : JC\MD
Sample : P3996-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Sep 18 01:52:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

