

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
 Data File : VN083953.D
 Acq On : 17 Sep 2024 18:17
 Operator : JC\MD
 Sample : P4003-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GM38-RW-3-MW-1-20240911

Quant Time: Sep 18 01:53:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	152800	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	263641	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	225611	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	95905	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	118916	51.613	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.220%			
35) Dibromofluoromethane	8.165	113	84530	50.083	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.160%			
50) Toluene-d8	10.565	98	319698	54.792	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 109.580%			
62) 4-Bromofluorobenzene	12.847	95	114769	47.675	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 95.360%			
Target Compounds						
44) Trichloroethene	9.347	130	24039	13.576	ug/l	98
64) Tetrachloroethene	11.106	164	1431	0.954	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
Data File : VN083953.D
Acq On : 17 Sep 2024 18:17
Operator : JC\MD
Sample : P4003-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GM38-RW-3-MW-1-20240911

Quant Time: Sep 18 01:53:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

