

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
 Data File : VN083959.D
 Acq On : 18 Sep 2024 03:42
 Operator : JC\MD
 Sample : VN0917MBL02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0917MBL02

Quant Time: Sep 18 05:43:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	148262	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	259116	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	219520	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	91781	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	115051	51.464	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.920%
35) Dibromofluoromethane	8.165	113	83968	50.619	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.240%
50) Toluene-d8	10.565	98	305867	53.337	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.680%
62) 4-Bromofluorobenzene	12.847	95	110296	46.617	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
Data File : VN083959.D
Acq On : 18 Sep 2024 03:42
Operator : JC\MD
Sample : VN0917MBL02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0917MBL02

Quant Time: Sep 18 05:43:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

