

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
 Data File : VN083962.D
 Acq On : 18 Sep 2024 05:09
 Operator : JC\MD
 Sample : P3985-01
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Sep 18 05:45:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 11 15:18:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	139038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	235685	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	196302	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	83270	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	107988	51.509	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.020%
35) Dibromofluoromethane	8.159	113	75516	50.050	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.100%
50) Toluene-d8	10.565	98	281137	53.898	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.800%
62) 4-Bromofluorobenzene	12.847	95	92714	43.082	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	86.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN091724\
Data File : VN083962.D
Acq On : 18 Sep 2024 05:09
Operator : JC\MD
Sample : P3985-01
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Sep 18 05:45:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N091024W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 11 15:18:56 2024
Response via : Initial Calibration

