

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091818\
 Data File : VN051297.D
 Acq On : 18 Sep 2018 15:01
 Operator : MD\SY
 Sample : PB112947TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112947TB

Quant Time: Sep 19 01:27:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 18 06:04:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	401288	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	609280	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	514885	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	142454	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	223311	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
35) Dibromofluoromethane	7.59	113	232508	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
50) Toluene-d8	10.09	98	814076	46.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.04%	
62) 4-Bromofluorobenzene	12.40	95	204549	36.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.50%	
Target Compounds						
16) Acetone	3.82	43	19134	10.309	ug/l	96
20) Methylene Chloride	4.55	84	23690	4.963	ug/l	99
43) Isopropyl Acetate	8.17	43	202005	34.432	ug/l	99
59) 2-Hexanone	10.77	43	13514	6.780	ug/l	66

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN091818\
Data File : VN051297.D
Acq On : 18 Sep 2018 15:01
Operator : MD\SY
Sample : PB112947TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112947TB

Quant Time: Sep 19 01:27:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 18 06:04:06 2018
Response via : Initial Calibration

