

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091818\
 Data File : VN051298.D
 Acq On : 18 Sep 2018 15:25
 Operator : MD\SY
 Sample : PB112947ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112947ZHE#01

Quant Time: Sep 19 01:28:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 18 06:04:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	403475	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	630556	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	518495	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	141419	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	226928	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
35) Dibromofluoromethane	7.59	113	236730	48.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.62%	
50) Toluene-d8	10.09	98	810906	44.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.54%	
62) 4-Bromofluorobenzene	12.40	95	202939	35.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.46%	
Target Compounds						
16) Acetone	3.82	43	19378	10.461	ug/l	99
20) Methylene Chloride	4.55	84	23622	4.922	ug/l	97
43) Isopropyl Acetate	8.17	43	220305	36.285	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN091818\
Data File : VN051298.D
Acq On : 18 Sep 2018 15:25
Operator : MD\SY
Sample : PB112947ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112947ZHE#01

Quant Time: Sep 19 01:28:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 18 06:04:06 2018
Response via : Initial Calibration

