

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091818\
 Data File : VN051301.D
 Acq On : 18 Sep 2018 16:39
 Operator : MD\SY
 Sample : J4945-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RB47195RT

Quant Time: Sep 19 01:41:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 18 06:04:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	400394	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	633516	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	506393	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	168055	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	223171	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
35) Dibromofluoromethane	7.59	113	233792	47.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.98%	
50) Toluene-d8	10.09	98	820740	45.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.20%	
62) 4-Bromofluorobenzene	12.40	95	237993	41.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.24%	
Target Compounds						
					Qvalue	
16) Acetone	3.82	43	89255	87.351	ug/l	100
20) Methylene Chloride	4.55	84	45798	9.616	ug/l	99
30) Chloroform	7.37	83	10187	1.287	ug/l	97
43) Isopropyl Acetate	8.17	43	158360	25.960	ug/l #	92
52) Toluene	10.16	92	8627	0.706	ug/l	96
59) 2-Hexanone	10.77	43	5138	2.479	ug/l	63
73) Isopropylbenzene	12.25	105	26305	1.856	ug/l #	72
95) Naphthalene	15.14	128	27313	10.502	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN091818\
Data File : VN051301.D
Acq On : 18 Sep 2018 16:39
Operator : MD\SY
Sample : J4945-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RB47195RT

Quant Time: Sep 19 01:41:16 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N091718W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 18 06:04:06 2018
Response via : Initial Calibration

